

PGT®130.DT - Calibration Unit PGT®130.DT (Part No.: 7100.PGT130.DT.CU)

- The Calibration Unit contains resistors for testing the Personnel Grounding Testers PGT®130.DT
- The Calibration Unit requires no battery or any external power supply
- Resistor values are selectable with a rotary switch
- A standard digital voltmeter is required for checking the PGT®130.DT test voltage

Technical Data

Setting	Value
100K	100.0 k Ω $\pm$ 0.5%
10M	10.0 M Ω $\pm$ 1%
20M	20.0 M Ω $\pm$ 1%
40M	40.0 M Ω $\pm$ 1%
100M	100.0 M Ω $\pm$ 1%
200M	200.0 M Ω $\pm$ 1%



Dimensions:
118 x 80 x 48 mm

- 1 year limited warranty
- Recommended calibration interval: 3 years